



Certificate of Conformance

Certificate Number: EXMXJ5473X123

Dated Issued: 07/29/2025

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-0000000000123
Expiration Date:	2027-07

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-24501A		
Processing Run Date:	07/25/2025		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.1
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.0

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company.
Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

Quality Department

Quality Manager

07/29/25

Date

7/30/25

Date

¹ Manufacturing data filed within the Quality Management System and available onsite.