



## *Certificate of Conformance*

Certificate Number: EXMXJ5485X58

Dated Issued: 08/13/2025

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.<sup>1</sup>

<b>Product Item Number:</b>	X101788
<b>Item Description:</b>	PACK, AMALGATOME SD 4" BLADE, 5 PC
<b>Item Lot Number:</b>	X-000000000058
<b>Expiration Date:</b>	07/18/2027

The blades were manufactured using a validated sterilization process:

<b>STERIS Gamma Process Run ID:</b>	1073-24758A		
<b>Processing Run Date:</b>	08/11/2025		
<b>Minimum Specified Dose (kGy):</b>	25.0	<b>Minimum Delivered Dose (kGy):</b>	27.3
<b>Maximum Specified Dose (kGy):</b>	45.0	<b>Maximum Delivered Dose (kGy):</b>	37.9


Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company.  
Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.



Quality Department

08/13/25

Date



Quality Manager

8/14/25

Date

<sup>1</sup> Manufacturing data filed within the Quality Management System and available onsite.