



Certificate of Conformance

Certificate Number: EXMXJ5542X60

Dated Issued: 10/21/2025

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101788
Item Description:	PACK, AMALGATOME SD 4" BLADE, 5 PC
Item Lot Number:	X-000000000060
Expiration Date:	09/10/2027

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-25678A		
Processing Run Date:	10/18/2025		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	26.3
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	37.1

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company.
Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

Release Cover

Quality Department



Quality Manager

10/21/25

Date

10/22/25

Date

¹ Manufacturing data filed within the Quality Management System and available onsite.