



Certificate of Conformance

Certificate Number: EXMXJ5625X140

Dated Issued: 12/11/2025

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-0000000000140
Expiration Date:	2027-11

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-26317A		
Processing Run Date:	12/06/2025		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.9
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	38.4

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**

Rebecca Coult
Quality Department

James E. Taniguchi
Quality Manager

12/11/2025
Date

12/11/25
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.