



Certificate of Conformance

Certificate Number: EXMXJ5635X141

Dated Issued: 12/31/2025

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-0000000000141
Expiration Date:	2027-11

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-26554A		
Processing Run Date:	12/28/2025 - 12/29/2025		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	32.2
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	42.7

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**

Rebecca Couer
Quality Department

Jessica Zwickler
Quality Manager

12/31/25
Date

31 Dec 2025
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.