



## *Certificate of Conformance*

Certificate Number: EXMXJ5667X67

Dated Issued: 12/31/2025

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.<sup>1</sup>

<b>Product Item Number:</b>	X101789
<b>Item Description:</b>	PACK, AMALGATOME SD 2" BLADE, 5 PC
<b>Item Lot Number:</b>	X-0000000000067
<b>Expiration Date:</b>	12/04/2027

The blades were manufactured using a validated sterilization process:

<b>STERIS Gamma Process Run ID:</b>	1073-26554A		
<b>Processing Run Date:</b>	12/28/2025 – 12/29/2025		
<b>Minimum Specified Dose (kGy):</b>	25.0	<b>Minimum Delivered Dose (kGy):</b>	32.2
<b>Maximum Specified Dose (kGy):</b>	45.0	<b>Maximum Delivered Dose (kGy):</b>	42.7

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company.  
Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

Releaser Cover

Quality Department

Jessica Zurcher  
Quality Manager

12/31/25

Date

31 Dec 2025

Date

<sup>1</sup> Manufacturing data filed within the Quality Management System and available onsite.