



Certificate of Conformance

Certificate Number: EXMJ5640X146

Dated Issued: 01/29/2026

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-0000000000146
Expiration Date:	2028/01

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-26919A		
Processing Run Date:	01/27/2026		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.0
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	38.0

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**

Rebecca Couer

Quality Department

James E. Taylor

Quality Manager

01/29/26

Date

1/29/26

Date

¹ Manufacturing data filed within the Quality Management System and available onsite.