



## Certificate of Conformance

Certificate Number: EXMJ5759X68

Dated Issued: 01/29/2026

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.<sup>1</sup>

<b>Product Item Number:</b>	X101789
<b>Item Description:</b>	PACK, AMALGATOME SD 2" BLADE, 5 PC
<b>Item Lot Number:</b>	X-0000000000068
<b>Expiration Date:</b>	01/09/2028

The blades were manufactured using a validated sterilization process:

<b>STERIS Gamma Process Run ID:</b>	1073-26919A		
<b>Processing Run Date:</b>	01/27/2026		
<b>Minimum Specified Dose (kGy):</b>	25.0	<b>Minimum Delivered Dose (kGy):</b>	27.0
<b>Maximum Specified Dose (kGy):</b>	45.0	<b>Maximum Delivered Dose (kGy):</b>	38.0

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**

*Rebecca Couer*  
Quality Department

*Jan E. Tang*  
Quality Manager

01/29/26  
Date

1/29/26  
Date

<sup>1</sup> Manufacturing data filed within the Quality Management System and available onsite.