



## Certificate of Conformance

Certificate Number: EXMJ5773X64

Dated Issued: 02/18/2026

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.<sup>1</sup>

<b>Product Item Number:</b>	X101788
<b>Item Description:</b>	PACK, AMALGATOME SD 4" BLADE, 5 PC
<b>Item Lot Number:</b>	X-0000000000064
<b>Expiration Date:</b>	02/02/2028

The blades were manufactured using a validated sterilization process:

<b>STERIS Gamma Process Run ID:</b>	1073-27201A		
<b>Processing Run Date:</b>	02/16/2026-02/17/2026		
<b>Minimum Specified Dose (kGy):</b>	25.0	<b>Minimum Delivered Dose (kGy):</b>	29.1
<b>Maximum Specified Dose (kGy):</b>	45.0	<b>Maximum Delivered Dose (kGy):</b>	36.8

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

*Rebecca Cover*  
Quality Department

*James E. T...*  
Quality Manager

02/18/2026  
Date

2/18/26  
Date

<sup>1</sup> Manufacturing data filed within the Quality Management System and available onsite.