



Certificate of Conformance

Certificate Number: EXMXJ5774X69

Dated Issued: 02/18/2026

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101789
Item Description:	PACK, AMALGATOME SD 2" BLADE, 5 PC
Item Lot Number:	X-0000000000069
Expiration Date:	02/02/2028

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-27201A		
Processing Run Date:	02/16/2026-02/17/2026		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	29.1
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	36.8

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

Releer Court
Quality Department

Jane E. Tamm
Quality Manager

02/18/26
Date

2/18/26
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.