Certificate of Conformance

Certificate Number: EXMXJ5776X70 Dated Issued: 02/18/2026

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.[1]

Product Item Number:	X101789
Item Description:	PACK, AMALGATOME SD 2" BLADE, 5 PC
Item Lot Number:	X-0000000000070
Expiration Date:	02/05/2028

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-27201A		
Processing Run Date:	02/16/2026-02/17/2026		
Minimum Specified Dose (kGy):	25.0	**Minimum Delivered Dose (kGy):**	29.1
Maximum Specified Dose (kGy):	45.0	**Maximum Delivered Dose (kGy):**	36.8

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company.
Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

Rebecca Cover
Quality Department

02/18/26
Date

James E. Tarquitt
Quality Manager

2/18/26
Date

[1] Manufacturing data filed within the Quality Management System and available onsite.