



Certificate of Conformance

Certificate Number: EXMXJ5812X150

Dated Issued: 03/13/2026

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-0000000000150
Expiration Date:	2028/02

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-27509A		
Processing Run Date:	03/11/2026		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	29.3
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	37.6

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**

Rebecca Cover
Quality Department

James E. Tappin
Quality Manager

03/13/26
Date

3/13/26
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.