



Certificate of Conformance

Certificate Number: EXMJ5822X154

Dated Issued: 03/26/2026

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-0000000000154
Expiration Date:	2028/03

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-27700A		
Processing Run Date:	03/24/2026		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.1
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	37.8

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

Rebecca Cover

Quality Department

Jessica Zwickler
Quality Manager

03/24/26

Date

26 mar 2026

Date

¹ Manufacturing data filed within the Quality Management System and available onsite.