



Certificate of Conformance

Certificate Number: EXMXJ5871X66

Dated Issued: 04/30/2025

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101788
Item Description:	PACK, AMALGATOME SD 4" BLADE, 5 PC
Item Lot Number:	X-00000000000066
Expiration Date:	04/12/2028

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-28117A		
Processing Run Date:	04/24/2026		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.0
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	37.3

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

Rebecca Coors
Quality Department

James E. Tappin
Quality Manager

04/30/26
Date

4/30/26
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.