



Certificate of Conformance

Certificate Number: EXMJ5891X72

Dated Issued: 05/13/2026

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101789
Item Description:	PACK, AMALGATOME SD 2" BLADE, 5 PC
Item Lot Number:	X-0000000000072
Expiration Date:	04/29/2028

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-28349A		
Processing Run Date:	05/09/2026		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.6
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	37.9

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

Rebecca Cover
Quality Department

James E. T... [Signature]
Quality Manager

05/13/20
Date

5/13/26
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.