



Certificate of Conformance

Certificate Number: EXMXJ5850X165

Dated Issued: 06/01/2026

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-0000000000165
Expiration Date:	2028/04

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-28576A		
Processing Run Date:	05/29/2026		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.1
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	37.9

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**

Rebecca Coover
Quality Department

James E. [Signature]
Quality Manager

06/01/26
Date

6/2/26
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.