



Certificate of Conformance

Certificate Number: EXMXJ5975X74

Dated Issued: 07/08/2026


The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101789
Item Description:	PACK, AMALGATOME SD 2" BLADE, 5 PC
Item Lot Number:	X-00000000000074
Expiration Date:	06/11/2028

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-29091A		
Processing Run Date:	07/04/2026		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	29.2
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.2

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company.
Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.


Quality Department


Quality Manager

07/08/26
Date

7/8/26
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.