



Certificate of Conformance

Certificate Number: EXMXJ5980X160

Dated Issued: 07/29/2026


The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹


Product Item Number:	X100457
Item Description:	ASSY, BOX OF 5 AMALGATOME MD PKG
Item Lot Number:	X-0000000000160
Expiration Date:	07/2028

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-29326A		
Processing Run Date:	07/25/2026		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.3
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	42.0

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company.
Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**


Quality Department


Quality Manager


Date


Date

¹ Manufacturing data filed within the Quality Management System and available onsite.