



Certificate of Conformance

Certificate Number: EXMXJ5983X76

Dated Issued: 07/29/2026

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101789		
Item Description:	PACK, AMALGATOME SD 2" BLADE, 5 PC		
Item Lot Number:	X-00000000000076		
Expiration Date:	07/07/2028		


The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-29326A		
Processing Run Date:	07/25/2026		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.3
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	42.0

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.


Quality Department

Quality Manager


Date

Date

¹ Manufacturing data filed within the Quality Management System and available onsite.