



Certificate of Conformance

Certificate Number: EXMXJ6006X174

Dated Issued: 08/17/2026



The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-00000000000174
Expiration Date:	2028/08

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-29555A		
Processing Run Date:	08/15/2026		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.2
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	37.3

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company.
Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.


 Quality Department

 Quality Manager

8/17/26
 Date
 8/18/26
 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.