Certificate of Conformance

Certificate Number: EXMXJ6048X71 Dated Issued: 09/28/2026

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.[1]

Product Item Number:	X101788
Item Description:	PACK, AMALGATOME SD 4" BLADE, 5 PC
Item Lot Number:	X-0000000000071
Expiration Date:	09/08/2028

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-30068A		
Processing Run Date:	09/23/2026		
Minimum Specified Dose (kGy):	25.0	**Minimum Delivered Dose (kGy):**	27.2
Maximum Specified Dose (kGy):	45.0	**Maximum Delivered Dose (kGy):**	36.8

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company.
Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

Quality Department — Date: 9/28/26

Quality Manager — Date: 9/28/26

[1] Manufacturing data filed within the Quality Management System and available onsite.