Certificate of Conformance

Certificate Number: EXMXJ6113X183 Dated Issued: 10/19/2026

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.[1]

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-0000000000183
Expiration Date:	2028/10

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-30360A		
Processing Run Date:	10/15/2026		
Minimum Specified Dose (kGy):	25.0	**Minimum Delivered Dose (kGy):**	30.9
Maximum Specified Dose (kGy):	45.0	**Maximum Delivered Dose (kGy):**	41.0

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company.
Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

Quality Department — Date: 10/19/26

Quality Manager — Date: 10/20/26

[1] Manufacturing data filed within the Quality Management System and available onsite.